

KIT PART #: VAIR-150REK FILTER ELEMENTS KIT



MOISTURE SEPARATOR

- › 5 micron filter
- › Removes moisture from compressed air to ensure dry air for respiratory breathing systems.



OIL FILTER ELEMENT

- › 0.01ppm coalescing filter
- › Filters out oil particles and vapors to prevent oil contamination in the breathing air.



CARBON FILTER ELEMENT

- › 0.003ppm carbon filter
- › Removes volatile organic compounds (VOCs) and odors to enhance breathing air quality



RECOMMENDATIONS

- › Replace filters every 3 months or when needle on differential gauge is on yellow during operations
- › De-pressure the unit before changing filters
- › Replace all 3 filters - do NOT replace individual filters

KIT PART #: CAL-1050 CALIBRATION KIT

The VAIR-150 is an autocalibration instrument - all calibration adjustments are made by the microprocessor using the calibration kit.



ZERO GAS BOTTLE



CARBON MONOXIDE BOTTLE



REGULATOR

RECOMMENDATIONS

- › Calibrate the unit every 90 days